

Versions: 01

Revision date: 25/11/2023

Pioneer Research Anahita Company No.157 Danesh street Technology Park Pardis Tehran/Iran

### 1. Identification

Product name: ELISA kits for Mouse of Mouse IL-33

Reactions: 96,48 rxns

Cat. No.: PRA-MIL-33 -96, PRA-MIL-33-α-48

### 2. Description

Interleukin-33 (IL-33), an inflammatory cytokine belonging to the interleukin-1 family, is predominantly produced by T helper type 2 (Th2) lymphocytes. Despite its association with various inflammatory properties and its receptor presence on numerous immune cells, IL-33 also plays a role in Th1-induced inflammation activation. Notably, this cytokine exhibits a protective function against arteriosclerosis. The current kit is meticulously designed and manufactured, utilizing mouse monoclonal anti-IL-33 antibodies. It's important to note that this kit is specifically tailored for use in measuring IL-33 levels in mouse models and may not be suitable for similar animal cases. The reported serum level of IL-33 in healthy mice typically falls within the range of 5 to 30 pg/ml. This information provides a baseline for understanding the physiological concentrations of IL-33 in a healthy murine context.

### 3. Kit Contents

Component	Cat. no	Quantity
Human anti-IL-33-α pre-	PRA-CIL-33 P	96 vials
coated plate		
Standards	PRA-IL-33 N1-4	200 μ1
HRP-Avidin buffer	PRA-HA	5 ml
HRP	PRA-HAA	540 μ1
Substrate	PRA-SU	5 ml
Stopping	PRA-ST	7 ml
10X washing buffer	PRA-WB	40 ml
Detection Ab	PRA- IL-33 D	5 ml







Versions: 01

Revision date: 25/11/2023

Pioneer Research Anahita Company No.157 Danesh street Technology Park Pardis Tehran/Iran

### 4. Storage specifications

All components of the ELISA kits can be stored at 4°C temperature.

### 5. Applications

Detection of both inflammatory and anti-inflammatory factors through ELISA.

# 6. Assay Procedure

How to Prepare Solutions:

- 1. Washing Buffer:
  - To prepare the washing solution, dilute it with distilled water at a 1:10 ratio.
- 2. HRP-Avidin:
- Spin the HRP vial using a microfuge device, then add all its contents to the HRP-Avidin buffer vial.
- For quantities less than 48 assays, mix 416  $\mu$ L of HRP-Avidin, 41  $\mu$ L of HRP per 8-well row.

Working with the Kit:

- 1. Plate Preparation:
- Remove the plate from its package and let it reach room temperature in a dry environment.
  - Add 50  $\mu L$  of standards #4 to #1 in the first to fourth wells.
- 2. Sample Incubation:
  - Add 50 microliters of the desired sample to the remaining wells.
  - Incubate for 60 minutes on a 200 RPM shaker at room temperature.
- 3. Plate Washing:
- After proper incubation, wash the plates three times using the washing solution (after adding the washing solution, incubate the plates for approximately 1 minute at room temperature and then drain).
- 4. Conjugated Antibody Addition:
  - Add 50 microliters of conjugated antibody (Detection ab) to all wells.







Versions: 01

Revision date: 25/11/2023

Pioneer Research Anahita Company No.157 Danesh street Technology Park Pardis Tehran/Iran

- Incubate for 60 minutes on a 200 RPM shaker at room temperature.
- 5. Plate Washing (Again):
  - After proper incubation, wash the plates three times using washing solution.
- 6. HRP-Avidin Addition:
  - Add 50 microliters of HRP-Avidin solution to all wells.
  - Incubate for 30 minutes on a shaker (at least at RPM 200).
- 7. Plate Washing (Again):
  - After proper incubation, wash the plates five times using washing solution.
- 8. Substrate Addition:
- Add 50 microliters of substrate to all wells and incubate for 15 minutes. Note that 15 minutes is enough for incubation, but if the amount of color produced is low, the time can be increased to 20 minutes.
- 9. Stopping Reaction:
  - Add 25 microliters of the stopping solution to all the wells.
- 10. Measurement:
- Measure the absorbance of the samples in an ELISA reader at a wavelength of 450 nm.

## 7. Safety

- The solutions used in the kit have oxidizing and acidic properties.
- Avoid direct contact with skin and eyes.
- In case of contact with the mentioned tissues, wash with plenty of water and go to the nearest medical center.

## 8. Quality Certifications

### 9. Further information

- This product is developed, designed, and sold exclusively only for research purposes use.
- The product was not tested for use in diagnostics or for drug development.
- It is not suitable for administration to humans or animals.









Versions: 01

Revision date: 25/11/2023

Pioneer Research Anahita Company No.157 Danesh street Technology Park Pardis Tehran/ Iran

### 10. Other Kits

#### Other ELISA kits:

#### Human:

IL-1 $\beta$ , IL-2, IL-4, IL-6, IL-8, IL-10, IL-12, IL-13, IL-18, IL-23, IL-29, IL-17A, TGF- $\beta$ , VEGF, TNF- $\alpha$ , IFN- $\gamma$ , CCL2 (MCP-1), CCL3 (MIP-1-alpha), CXCL10 (IP-10), CXCL12 (SDF-1), CCL21

#### Mouse:

IL-1β, IL-2, IL-4, IL-6, IL-10, IL-13, IL-33, IL-18, TNF-α, TGF-β, CCL3, IFN-γ, Total IgG, Total IgE

#### Rat:

TNF-α, IL-1β, IL-6, IL-10, IL-17A

#### **NOTE**

All products have been manufactured by Karmania Pars Gene Company in the Iran.

